

Test Certificate

regarding
Environmental and Functional Performance Requirements



CEcert GmbH, Alter Holzhafen 19a, 23966 Wismar (Germany)
Test Laboratory

Certificate No.: 613.225.2

License Holder:	Dräger Safety AG & Co. KGaA Revalstraße 1 23560 Lübeck, Germany
Product identification::	Breath alcohol test device
Type/Model/Parameters:	Alcotest 6820
P/N; S/N:	P/N: 8322620 S/N: AREC-0235 used for tests according to chapter 7.8 Electrical disturbances S/N; AREC-0227 used for tests according to chapter 7.4.2 Accuracy testing; 7.4.3 Repeatability testing; 7.5 Memory effects 7.6.3 Ambient relative humidity; 7.6.4 Interfering substances; 7.6.5.3 Influence of volume and time during the breath exhalation; 7.6.5.4 Influence of flow rate and time during the breath exhalation; 7.6.6 Voltage variation; 7.7 Mechanical and climatic disturbances S/N: AREC-0200; AREC-0337 used for tests according to chapter 7.4.4 Drift testing S/N: AREC-0205; AREC-0332 used for tests according to chapter 7.6.2 Operating temperature S/N: AREC-0203 used for tests according to chapter 7.6.5.2 Minimum volume test; 7.6.5.5 Influence of variation of alcohol concentration during the breath exhalation; 7.6.5.6 Influence of pressure and flow rate during the breath exhalation; 7.6.7 Power supply duration
Applied standards:	EN 15964:2011 Increased operating temperature range -5°C...+50°C te sted
Restrictions:	none
Classifications	Power supply: Internal battery 2x1,5V AA Drift testing: 6 Month
Test report(s):	#413225.1; #413225.2; #413225.3; #413225.4
Test result:	PASS

This test certificate is relevant exclusively to the item(s) submitted for testing.
Consider the test report specified in this certificate and the safety indications specified in the technical documentation to the product.

Approved by



Wismar, Tuesday, 28 January 2014

Dipl.-Ing. Bernd Schmidt